


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AU-155C-C-P-0P1-T1PA-L23*

S.N.: *7927*

Date: *MAR 14 1997*



- a) Output signal amplitude:
PULSE: 0 TO +2.0 AMP
DC: 0 TO +500 mA.
- b) Pulse width:
100 NS TO 10 US
- c) Rise time:
≤ 50 NS
- d) Fall time:
≤ 50 NS
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability:
OK
- g) Prime power: *120/240V*
50-60 Hz.

TOP MONITOR OUT TO 50Ω (500mV/DIV)
DOT LOAD VOLTAGE (1V/DIV)

