


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-2-C-P-M-CA-OW - BHULA

S.N.: 7930

Date: MAR 10 1997



- a) Output signal amplitude:
0 TO +350 V
- b) Pulse width: (50 n (0.05 μs))
5 TO 100 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 3 NS
- e) PRF: 0 TO 5 KHz.
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz.

70 db ATTN
- - ≈ 320 VOLTS/DIV
PRF ≈ 1.0 KHz

[Handwritten signature]