


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-3A-C-P*

S.N.: *7939*

Date: *APRIL 3 1997*



- a) Output signal amplitude: *725 to 775 AMP*
- b) Pulse width: *≤ 2.5 ns*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 to 20 kHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*  
*50-60 Hz*

*f<sub>c</sub> = 5.0 n*  
*≈ 100 V/Div*  
*∴ ≈ 20 AMP/Div*  
*PRF = 20 kHz*

*✓*  
*[Signature]*