


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-BA-C-P

S.N.: 7948

Date: APRIL 3 1997



- a) Output signal amplitude:
0 TO +40 VOLTS TO
- b) Pulse width: $R_L \geq 0.3 \Omega$ (50 AMP. MAX)
2 μ S TO 20 μ S
- c) Rise time:
 $\leq 0.5 \mu$ S
- d) Fall time:
 $\leq 0.5 \mu$ S
- e) PRF: 0.1 KHz TO 1.0 KHz
(100% DUTY CYCLE)
- f) Jitter, stability:
O/L
- g) Prime power:

a) 120/240 V
50-60 Hz

b) 0 TO +40 VDC,
50 AMPS MAX

$R_L = 0.2 \Omega$ MAX
VOLT: 5 VOLTS/DIV
(25 AMPS/DIV)
TIME: 50 μ S/DIV