


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVP-AN-1N2-C-P-M*

S.N.: *7953*

Date: *MAR 31 1997*



a) Output signal amplitude:  
*0 TO + 20 VOLTS*

b) Pulse width:  
*0.3 TO 2.0 NS*

c) Rise time:  
*≤ 100 PS*

d) Fall time:  
*≤ 200 PS*

e) PRF:  
*0 TO 1.0 MHz*

f) Jitter, stability:  
*OK*

g) Prime power:  
*120/240V*  
*50-60 Hz.*

*40 db ATTEN*  
*10 VOLTS/DIV*  
*TOP 1 NS/DIV*  
*BOT 200 PS/DIV*  
*PRF = 1.0 MHz*