


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMH-1-P*

S.N.: *7959*

Date: *FEB 4 1997*



40 dB ATTN

10 VOLT/DIV

PRF ≈ 10 MHz

- a) Output signal amplitude: *0 TO +10V*
- b) Pulse width: *≤ 200ps*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *400 mA MAX*

+24 VDC

400 mA MAX

[Signature]