


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM4-1P*

S.N.: *7961*

Date: *FEB 4 1997*



- a) Output signal amplitude: *0 to +10V*
- b) Pulse width: *520ps*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 to 25 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+24 VDC*

40 dB ATTEN
10 V/8 DIV
PRF = 10 MHz

400 mA (MAX)
[Signature]