


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVMT4-1-P*
S.N.: *7962*
Date: *FEB 4 1997*



- a) Output signal amplitude: *0 TO +70V*
- b) Pulse width: *≤ 200 ps*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+24 VDC*

46 dB ATTEN
- 10 V0 V3/DIV
PRF ≈ 25 MHz

400 mA SMIA
[Signature]