


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ANM4-1-P*

S.N.: *7964*

Date: *FEB 5 1997*



- a) Output signal amplitude: *0 to +10V*
- b) Pulse width: *≤ 20ns*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *10 to 25 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+24 VDC*

*40 db atten  
10 VOLTS/Div  
PRF = 25 MHz.*

*+24 VDC*

*400mA (approx)  
[Signature]*