


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-1-CP*

S.N.: *7973*

Date: *MAR 24 1997*



- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
0.2 TO 4.0ns
- c) Rise time:
 $\leq 100ps$
- d) Fall time:
 $\leq 135ps$
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
100 ± 10% VOLT
50-60 MHz
[Signature]