


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANMP-2-C-P-DBM1*

S.N.: *7974*

Date: *APRIL 29 1997*



- a) Output signal amplitude: *0 TO +15 VOLTS*
- b) Pulse width: *($V_c = 50\Omega$)
5 TO 100 NS*
- c) Rise time: *≤ 140 PS*
- d) Fall time: *≤ 150 PS*
- e) PRF: *0 TO 1 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V
50/160Hz*

40 db ATTEN
- 10 VOLTS/DIV
TOP 10 NS/DIV
MMW 200PS/DIV (RISE TIME)
BOT 200PS/DIV (FALL TIME)
PRF = 100 KHz

[Signature]