


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-2-e-P
S.N.: 8000
Date: July 4 1994



- a) Output signal amplitude: 0 TO +350 VOLTS
- b) Pulse width: $t_c = 50\text{ns}$
5 TO 100 NS
- c) Rise time: $\leq 2\text{ NS}$
- d) Fall time: $\leq 3\text{ NS}$
- e) PRF: 0 TO 5 KHz
- f) Jitter, stability: OK
- g) Prime power:

70 dB ATTEN
 $\approx 320\text{ VOLTS/DIV.}$
PRF $\approx 1\text{ KHz.}$

120/240V
50-60 Hz