


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-7B-C-N-R5*

S.N.: *3009*

Date: *AUG 16 1994*



R_L = 50 Ω
2 μ s / DIV.

- a) Output signal amplitude: *0 TO -700 VOLTS*
- b) Pulse width: *R_L \geq 50 Ω*
- c) Rise time: *0.1 TO 100ns*
(0.5% MAX DUTY CYCLE)
 \leq 50 ns
- d) Fall time: *\leq 50 ns*
- e) PRF: *0 TO 10 KHz*
(0.5% MAX DUTY CYCLE)
- f) Jitter, stability: *OK*
- g) Prime power: *120 / 240 V*
50 - 60 Hz

[Signature]