


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ADR-3-PS-CP*


S.N.: *8015*

Date: *JULY 29 94*



- a) Output signal amplitude:
0 to +20V
- b) Pulse width: *R_L ≥ 50 Ohms*
0.1 μs to 100 μs
- c) Rise time: *(0.5% max duty cycle)*
≤ 10 ns
- d) Fall time:
≤ 10 ns
- e) PRF: *0 to 10 kHz*
(0.5% max duty cycle)
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
52-60 Hz

① *R_L = 50 Ω PRF = 2 kHz*



② *R_L = 50 Ω PRF = 50 Hz*