


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MV-155A-44mA*


S.N.: *8033*

Date: *APR 16 1964*




- a) Output signal amplitude:
0 TO -170 mA
- b) Pulse width:
= 20 NS TO DC
- c) Rise time:
≤ 10 NS
- d) Fall time:
≤ 10 NS
- e) PRF:
DC TO 40 MHz
- f) Jitter, stability:
OK
- g) Prime power:
-15 VDC, 200 mA
+ 5.8 VDC, 40 mA

Ⓐ $V_{in} = 0.3 V$
 $I_{out} = 55 mA$



Ⓑ $V_{in} = 10 V$
 $I_{out} = 170 mA$
PRF ≈ 500 kHz



RISE TIME
FALL TIME
CHECK

