


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-3C-C-P*

S.N.: *3045*

Date: *SEPT 20 1994*



- a) Output signal amplitude: *0 TO +120 VOLTS*
- b) Pulse width: *(TO $R_L = 50 \Omega$)*
- c) Rise time: *TO 100 NS*
≤ 300 PS
- d) Fall time: *≤ 2 NS*
- e) PRF: *0 TO 5 KHz.*
- f) Jitter, stability: *OK*
- g) Prime power: *120 / 240 VOLTS*
50-60 KHz.

60 db ATTEN
100 VOLTS/DIV
TOP: 10 NS/DIV
BOT: 1 NS/DIV
PRF = 5 KHz
 $R_L = 50 \Omega$

[Signature]