


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-3A-PS-P*  
S.N.: *8046*  
Date: *AUG 19 1994*



- a) Output signal amplitude: *25 TO 75 AMPS*
- b) Pulse width:  *$\leq 2.5 \text{ ns}$*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 20 kHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120 / 240 V*  
*50 - 60 kHz*

*$R_L = 5.1 \Omega$*   
 *$\approx 100 \text{ VOLTS/DIV}$*   
 *$\approx 20 \text{ AMPS/DIV}$*   
*PRF  $\approx 10 \text{ kHz}$*

*120 / 240 V*  
*50 - 60 kHz*  
*JD*