


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-2-PS-P-SHW*

S.N.: *8060*

Date: *Jul 15 1994*



70. dB ATTEN
= 320 VOLTS/DIV.
PRF = 1.0 KHz

- a) Output signal amplitude:
0 TO +350 VOLTS
- b) Pulse width: *($T_0/2 = 50ns$)*
- c) Rise time:
5 TO 100 NS
≤ 2 NS
- d) Fall time:
≤ 3 NS
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50 - 80 Hz
AA