


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVI-V-C-P-PN

S.N.: 8065

Date: AUG 29 '94



- a) Output signal amplitude:
0 TO ± 50 VOLTS
- b) Pulse width:
2 TO 100 NS
- c) Rise time:
 ≤ 350 PS
- d) Fall time:
 ≤ 2 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 VOLTS
50-60 Hz

50 db ATTEN
 ≈ 32 VOLTS / DIV
TOP 5 NS / DIV
BOT 500 PS / DIV
(RISE TIME)

PRF = 20 KHz

[Signature]