


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-M-1413-C-05-P-0541**

S.N.: *8077*

Date: *OCT 5 1994*



- a) Output signal amplitude:
0 TO +40 VOLTS
- b) Pulse width:
0.4 TO 2.0 NS
- c) Rise time:
≤ 150 PS
- d) Fall time:
≤ 250 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 VOLTS

50 dB ATTEN
32 VOLTS DIV
PRF = 100 KHz

50-60 MHz

** -0541 SUFFIX DENOTES A -3 VOLT TRIG OUT PULSE.*