


PULSE GENERATOR
PERFORMANCE CHECK

Model: MO-9A-C-P

S.N.: 3031

Date: JUN 14 1997



- a) Output signal amplitude:
0 TO +200mA
- b) Pulse width:
0.4 TO 4.0NS
- c) Rise time:
≤ 200ps
- d) Fall time:
≤ 200ps
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60Hz

TOP MAINFRAME OUT
TO 40db
10V/DIV
1NS/DIV

BUT ME OUT TO 20db
PRF ≈ 100KHz

[Signature]