


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVIR-2-P*

S.N.: *8086*

Date: *MAY 16 1996*



- a) Output signal amplitude:  
*0 TO +200 V*
- b) Pulse width: *CTO 50ns*
- c) Rise time:  
*2 TO 10 NS*  
*≤ 1 NS*
- d) Fall time:  
*≤ 2 NS*
- e) PRF: *0 TO 50 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15 V DC*  
*400 mA (MAX)*

*60 db ATTEN*  
*- 100 VOLTS/DIV*  
*5 NS/DIV*  
*PRF = 20 KHz.*