


PULSE GENERATOR
PERFORMANCE CHECK

Model: MR-3-PW-PS-PW-NUM B


S.N.: 8038

Date: JUNE 17 1997



- a) Output signal amplitude:
0 TO ± 200 VOLTS
- b) Pulse width: $R_L = 50 \Omega$
0.1 TO 100 μ S.
(0.5% MAX DUTY CYCLE)
- c) Rise time:
 ≤ 10 NS
- d) Fall time:
 ≤ 10 NS

Ⓐ Fast, $R_L = 50 \Omega$ PRF = 1 kHz.



- e) PRF: 0 TO 10 KHz
(0.5% MAX DUTY CYCLE)
- f) Jitter, stability:
O.K.
- g) Prime power: 120/250 V
50 60 Hz

[Handwritten signature]

Ⓞ Normal, 50 V/DIV,
10 ns/DIV
 $R_L = 50 \Omega$
PRF = 50 Hz