

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-C-P*
 S.N.: *8094*
 Date: *APRIL 29 1997*



- a) Output signal amplitude: *0 TO +15 VOLTS*
- b) Pulse width: *(50%)*
0.2 TO 2.0 NS
- c) Rise time: *≤ 100 PS*
- d) Fall time: *≤ 135 PS*
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50 to 10A

40 dB ATTEN

10 V/DIV

TDR 1 NS/DIV
PRF = 12 MHz

BDI 5 NS/DIV
PRF = 25 MHz

[Signature]