


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1-F*

S.N.: *2104*

Date: *APRIL 16 1997*



- a) Output signal amplitude:
+25 to +100 amps
- b) Pulse width:
≤ 25 ns (1/2 power)
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 to 1 kHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 VDC
≈ 200 mA.

≈ 60 AMP/DIV

PRF = 10 kHz

[Signature]