


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-MFP-P-UFA*

S.N.: *3110*

Date: *JUNE 12 1997*



- a) Output signal amplitude:
 $\geq +40$ VOLTS
- b) Pulse width:
1 TO 100 NS
- c) Rise time:
 ≤ 350 PS
- d) Fall time:
 ≤ 2 NS
- e) PRF:
MIN PW: 0 TO 1 MHz
MAX PW: 0 TO 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 VDC
100 mA
(MAX)

50 dB ATTEN
 $\therefore 32$ VOLTS/PW
PRF = 30 kHz

[Signature]