


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVI-MP-P-URA

S.N.: 8112

Date: JUNE 12, 1997



- a) Output signal amplitude:
 $\geq + 40 \text{ VOLTS}$
- b) Pulse width:
 $1 \text{ TO } 100 \text{ NS}$
- c) Rise time:
 $\leq 350 \text{ PS}$
- d) Fall time:
 $\leq 2 \text{ NS}$
- e) PRF:
MIN PW: 0 TO 1 MHz
MAX PW: 0 TO 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VDC
100 mA (MAX)

50 dB ATTEN

$\therefore 32 \text{ VOLTS/DIV}$

PRF = 30 kHz