


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVI-MP-P-URA

S.N.: 8113

Date: JUNE 12, 1997



- a) Output signal amplitude:
 $\geq +40$ VOLTS
- b) Pulse width:
1 TO 100 NS
- c) Rise time:
 ≤ 350 PS
- d) Fall time:
 ≤ 2 NS
- e) PRF:
MIN PW: 0 TO 1 MHz
MAX PW: 0 TO 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VDC
100 mA (MAX)

50 dB ATTEN
 $\therefore 32$ VOLTS/DIV
PRF = 30 kHz