


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVI-MP-P-URA

S.N.: 8116

Date: JUNE 12, 1997



- a) Output signal amplitude:  
 $\geq +40$  VOLTS
- b) Pulse width:  
1 TO 100 NS
- c) Rise time:  
 $\leq 350$  PS
- d) Fall time:  
 $\leq 2$  NS
- e) PRF:  
MIN PW: 0 TO 1 MHz  
MAX PW: 0 TO 20 kHz
- f) Jitter, stability:  
OK
- g) Prime power:  
+15 VDC  
100 mA (MAX)

50 dB ATTEN

$\therefore$  32 VOLTS/DIV

PRF = 30 kHz