


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVRF-4-C-P

S.N.: 3121

Date: AUG 5 1997



- a) Output signal amplitude: 0 TO +400 VOLTS
- b) Pulse width: ( $R_L = 50 \Omega$ ) 0.1 TO 10  $\mu$ S
- c) Rise time:  $\leq 5$  NS
- d) Fall time:  $\leq 5$  NS
- e) PRF: 0 TO 1 KHz
- f) Jitter, stability: O.K.

$R_L = 50 \Omega$   
100 NS/DIV  
 $\approx 150$  VOLTS/DIV  
PRF = 1 KHz

- g) Prime power: 120/240 V  
50-60 Hz

RISE TIME  
60 db ATTEN  
100 VOLTS/DIV  
10 NS/DIV  
PRF = 1 KHz

