


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV6-4B-C-N-CA*

S.N.: *8127*

Date: *JUNE 11 1997*



- a) Output signal amplitude: *0 TO -850 V*
- b) Pulse width: *(RL=50Ω)*
≤ 6 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 3.5 NS
- e) PRF: *0 TO 10 KHz*
- f) Jitter, stability:
OK
- g) Prime power: *120/240V*
50-60Hz

70 dB ATTN
≈ 320 VOLTS/DIV
PRF = 2 kHz

[Signature]