


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVP-M-1-N-C-3M1

S.N.: 8138

Date: JUN 29 1997



- a) Output signal amplitude: 0 TO 10 V
- b) Pulse width: 1 TO 10 NS
- c) Rise time: ≤ 100 PS
- d) Fall time: ≤ 200 PS
- e) PRF: 0 TO 200 KHz
- f) Jitter, stability: OK
- g) Prime power: 170/240V
50-60Hz

40 dB ATTN
10 VOLTS/DIV
TOP 2 NS/DIV
BOT 100 PS/DIV
(RISE TIME)