


PULSE GENERATOR
PERFORMANCE CHECK

Model: *NL-3C-C-P*

S.N.: *8139*

Date: *June 26 1991*



- a) Output signal amplitude: *0 to +130 Volts*
- b) Pulse width: *(to 50%)*
- c) Rise time: *3 to 100ns*
≤ 300ps
- d) Fall time: *≤ 2ns*
- e) PRF: *0 to 5kHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50 dBm

60 db return
-100 VdBm DN
TRR 10 ns/DIV
DOT 1 ns/DIV
(RISETIME)

PRF = 5kHz

$R_L = 50\Omega$

[Signature]