


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MH-5-1-P*

S.N.: *8161*

Date: *Aug 15 1997*



- a) Output signal amplitude: *0 TO +10V*
- b) Pulse width: *≤ 130 ns*
- c) Rise time: *—*
- d) Fall time: *—*
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power:

*+15 VDC*  
*300 mA (max)*

*40 dB ATTE*

*.10 VOLTS / DIV*

*TOP: 2 NS / DIV*

*BOT: 200 NS / DIV*

*PRF = 300 kHz*

*[Signature]*