


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *M14-P*  
S.N.: *B164*  
Date: *Aug 10 1997*



- a) Output signal amplitude: *0 TO +30 VOLTS*
- b) Pulse width: *≤ 1 NS*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 1 pulse*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 VDC*  
*200 mA (MAX)*

*40 dB ATTEN*  
*10 VOLTS/DIV*  
*2 NS/DIV*  
*PRF = 300KHz*

*[Signature]*