


PULSE GENERATOR  
PERFORMANCE CHECK

Model: APP-2-PS POT


S.N.: 8178

Date: Aug 6 1997

- a) Output signal amplitude:  
0 to +20 VOUT
- b) Pulse width:  
0.4 to 100 ns
- c) Rise time:  
≤ 200 ps
- d) Fall time:  
≤ 300 ps
- e) PRF:  
0 to 100 kHz
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240V  
50-60Hz



A) 8 NS RANGE  
10 V/DIV, 1 NS/DIV



B) 100 NS RANGE  
10V/DV  
TOP: 2 NS/DIV  
BOT: 10 NS/DIV

*[Handwritten signature]*