


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVIR-3-C-P-M1

S.N.: 8198

Date: OCT 8 1997



a) Output signal amplitude:

0 TO +125V

b) Pulse width:

10 TO 50 NS

c) Rise time:

≤ 2 NS

d) Fall time:

≤ 2 NS

e) PRF:

0 TO 50 KHz

f) Jitter, stability:

OK

g) Prime power:

100 ± 10 W0dB

50-60 Hz

60 db ATTN

100 W0dB 10W

PRF = 50 KHz