


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMMN-1-P-PS*

S.N.: *8213*

Date: *SEPT 25 1997*



- a) Output signal amplitude: *0 TO +5V*
- b) Pulse width: *0.3 TO 5.0N*
- c) Rise time: *≤ 100 PS*
- d) Fall time: *≤ 200 PS*
- e) PRF: *0 TO 50 MHz*
- f) Jitter, stability: *OK*

40 db ATTEN
5 VOLTS/DIV
PRF ≈ 30 MHz

g) Prime power: *120/240V*
50-60Hz

[Signature]