


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR14-3-PS-PN*

S.N.: *8223*

Date: *OCT 9 1997*



- a) Output signal amplitude: *0 TO ±3000 VOLTS*
- b) Pulse width: *($R_L \ge 10K$)*
- c) Rise time: *200 NS TO 2.5 μ S*
- d) Fall time: *$\le 100 NS$*
- e) PRF: *$\le 100 NS$*
- f) Jitter, stability: *OK*

g) Prime power: *120/240V*
50-60Hz

$R_L = 10K$
500 NS/DIV
500 V/DIV
PRF = 1KHz

[Signature]

