


PULSE GENERATOR
PERFORMANCE CHECK

Model: MR-F01-PS

S.N.: 8241

Date: JAN 8 1998



- a) Output signal amplitude:
TTL, $R_L \geq 50\Omega$
- b) Pulse width:
50NS TO 50US
- c) DIVISION FACTOR:
2 TO 999
- d)
- e) PRF: 0 TO 10 MHz
- f) Jitter, stability: OK
- g) Prime power: 120/240V
50-60 Hz

TTL output, $R_L = 50\Omega$
 $\div 10$

BOI INPUT ≈ 10 MHz