


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AA-51-ME-MU-MD-KLA2*


S.N.: *3242*

Date: *SEPT 22 1997*



- a) Output signal amplitude: *NA.*
- b) Pulse width: *0.4 TO 200ns*
- c) Rise time: *≤ 200ps*
- d) Fall time: *≤ 200ps*
- e) PRF: *0 TO 1 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *NA*

*(A) TPI OUT TO 20db
(DRIVEN BY AVD-9A-C-N)*



*(B) DETECTED OPTICAL OUTPUT
DETECTOR: MDT MARD 500
TIA: ANTECH AU-149-BW4-2.8K
-PIN-SP2.*