


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AM-3A-PW-C*

S.N.: *8247*

Date: *DEC 1 1997*



- a) Output signal amplitude: *0 TO +200V (RL ≥ 50Ω)*
- b) Pulse width: *50NS TO 5.0μS*
- c) Rise time: *(0.5% MAX PULSE CYCLE)*  
*≤ 10NS*
- d) Fall time: *≤ 10NS*
- e) PRF: *0 TO 10kHz*  
*(0.5% MAX OUTPUT CUR)*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*  
*50-60Hz*

*[Signature]*

*TDR (50V/DIV) 1kHz*  
*(145/DIV)*

*DOE (50V/DIV) 10kHz*  
*(100NS/DIV)*  
*Rc = 50Ω*