


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVIR-2-P

S.N.: 8255

Date: NOV 17 1997



- a) Output signal amplitude: 0 TO +200V (TO 50V)
- b) Pulse width: 2 TO 10 NS
- c) Rise time: < 1 NS
- d) Fall time: < 2 NS
- e) PRF: 0 TO 50 KHz
- f) Jitter, stability: OK
- g) Prime power: +15 VDC.  
400 mA (MAX)

60 db ATTN  
100 VOLTS/DIV  
5 NS/DIV  
PRF = 10 KHz