


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-GI-C-PN-MITCI-0P1*


S.N.: *8265*

Date: *DEC 16 1997*



- a) Output signal amplitude: *0 TO ± 200 VOUT*
- b) Pulse width: *(R_L > 10K) 100 NS TO 100 μS*
- c) Rise time: *≤ 20 NS*
- d) Fall time: *≤ 20 NS*
- e) PRF: *0 TO 1 KHz*

*(A) Load R_L = 50K
20 μS/DIV
50 V/DIV*



- f) Jitter, stability: *OK*
- g) Prime power: *120/240V
50-60 Hz*

[Handwritten signature]

*Next R_L = 100K
100 NS/DIV
50V/DIV*