


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMP-3-P*

S.N.: *8277*

Date: *OCT 27 1997*



- a) Output signal amplitude: *0 to +20V*
- b) Pulse width: *(to 50%)*  
*8 to 100ns*
- c) Rise time: *≤ 300ps*
- d) Fall time: *≤ 300ps*
- e) PRF: *0 to 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+24VDC*  
*400 mA*  
*mA*

*40 dB ATTEN*  
*10 VOLTS / DIV*  
*TOP 10 NS / DIV*  
*BOT 1 NS / DIV*  
*(RISE TIME)*

*[Signature]*