


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANMP-3-P*
S.N.: *8278*
Date: *OCT 27 97*



- a) Output signal amplitude: *0 TO +20 VOLTS*
 - b) Pulse width: *(TO 50N)*
8 TO 100 N)
 - c) Rise time: *≤ 300 ps*
 - d) Fall time: *≤ 300 ps*
 - e) PRF: *0 TO 1.0 MHz.*
 - f) Jitter, stability: *OK*
 - g) Prime power: *+24 VDC*
400 mA (MAX)
- AB*

40db ATTN
10 VOLTS/DIV

TOP 10 NS/DIV

BOT 1 NS/DIV
(RISE TIME)