


PULSE GENERATOR
PERFORMANCE CHECK

Model: AO-7A-P5-P

S.N.: 8288

Date: JAN 13 1998



- a) Output signal amplitude: 0 TO +20 VOLTS TO
- b) Pulse width: $R_L \geq 0.5 \Omega$
(40 AMP MAX)
- c) Rise time: 0.2 TO 200 NS (1% MAX DUTY CYCL)
 $\leq 100 \text{ NS}$
- d) Fall time: $\leq 100 \text{ NS}$
- e) PRF: 0 TO 1 KHZ
(1% MAX DUTY CYCL)
- f) Jitter, stability: OK

- g) Prime power: 120/240V
50-60 HZ
- ly

$R_L = 0.5 \Omega$
2 us/DIV
200 NS/DIV
(RISE TIME)