


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MV-1011A-C*



S.N.: *B291*

Date: *NOV 25 1997*



- a) Output signal amplitude: *0 TO ± 100 VOLTS*
- b) Pulse width: *(R<sub>L</sub> = 50 Ω)  
100 NS TO 1 μS*
- c) Rise time: *(10% MAX DUTY CYCLE)  
≤ 10 NS*
- d) Fall time: *≤ 10 NS*
- e) PRF: *0 TO 1 MHz  
(50% MAX DUTY CYCLE)*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V  
50-60 Hz*

Ⓐ *Foot.*



Ⓑ

*Next R<sub>L</sub> = 50 Ω, R<sub>out</sub> = 50 Ω  
10V/DIV, 100 NS/DIV  
PRF ≈ 1 kHz*

*MS*