


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-7A-C-N-0P1B*

S.N.: *8295*

Date: *JAN 22 1998*



- a) Output signal amplitude: *0 TO -700 VOLTS TO*
- b) Pulse width: *$R_L \geq 50 \Omega$
50 NS TO 5.0 US*
- c) Rise time: *(0.5% MAX DUTY CYCLE)
 $\leq 50 \text{ NS}$*
- d) Fall time: *$\leq 50 \text{ NS}$*
- e) PRF: *0 TO 10 KHz
(0.5% MAX DUTY CYCLE)*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V
50-60 Hz*

*150V/DIV
200NS/DIV
PRF = 1KHz
 $R_L \geq 50 \Omega$*