


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVO-3B-C-F

S.N.: 8301

Date: FEB 4 1998



- a) Output signal amplitude:  
+10 TO +30 AMP
- b) Pulse width:  
1.5 NS (HALF POWER)
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 20 KHz
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240V,  
50-60 Hz

$R_L = 5.0 \Omega$

20 AMP / DIU

PRF = 20 KHz

WJ